



Vijayalakshmi
Health & Surgicals Pvt. Ltd.

Sterile Latex Surgical Gloves Powder-Free Product Catalogue

Abstract

Catalogue describes about the company and product specifications of latex surgical gloves powder-Free



Vijayalakshmi Health & Surgicals Pvt. Ltd.
406, APIIC Growth Centre, Gundlapalli,
Ongole, Prakasam Dt – 523211,
Andhra Pradesh, India.
info@vlhsglove.com

About Company

Vijayalakshmi Health & Surgicals (VLHS) Pvt Ltd is one of the India's leading manufacturers of good quality Latex Surgical gloves and Latex Examination gloves to protect health care professionals, workers, and patients. VLHS gives priority to keep your patients and surgical staff safe from cross contamination, infection, and allergic reaction. Company's undeterred commitment to provide the best quality products and services to customers by maintaining ASTM and EN455 standards enabled us to become pioneer in Glove manufacturing.

Ours State-of-the-art facility is equipped by latest fully automated double former machinery and globally recognized operations standards. VLHS is standing upon a strong infrastructural base, continually striving for improvements, innovations to set new standards in Healthcare Industry and serve customers beyond expectations.

PRODUCT DESCRIPTION

Surgical Gloves manufacture by VLHS are well designed for surgeons featuring the outstanding fit, comfort, dexterity, flexibility, and protection. Polymer coated powder free interior enables incredible smooth donning experience.

The sterile powder free (polymer coated) latex surgical gloves are disposable device made up of natural rubber latex. Gloves provide barrier protection to both patients and health care workers against exchanging microorganisms through their hands and body fluids. Because of the excellent elasticity, durability, tactile properties, low microbial/fluid penetration rate and low cost, natural rubber latex (NRL) gloves have been most widely used. Surgical gloves have more precise sizing with a better precision and sensitivity and are made to a higher standard.

The gloves are designed and available in specified sizes as of anatomically shaped with polymer coating. The sterile latex surgical gloves appear in natural white color, micro rough textured in the palm area, highly elastic and tear resistant. Product is classified under Medical device Class IIa and Personal protective equipment category III.



UDI-DI Code : 890615299LSG-PF-E7Z

REF No	Size	Pack qty
0117	5.5	50 pairs/box
0118	6	50 pairs/box
0119	6.5	50 pairs/box
0120	7	50 pairs/box
0121	7.5	50 pairs/box
0122	8	50 pairs/box
0123	8.5	50 pairs/box
0124	9	50 pairs/box

APPLICATION / INTENDED USE

The sterile latex surgical gloves are intended to prevent contamination of the patient during invasive procedures and the user to protect their hand from exposure to potentially infectious materials.

DEVICE DETAILS

PRODUCT NAME:	Sterile Latex Surgical Gloves
BRAND/PROPRIETARY NAME:	DR.GLOVE
MODEL/VARIANTS:	Powder Free
EMDN CODE:	T01010102 - Surgical Gloves, Latex, Non-Powdered
MDA/MDN CODE:	MDN 1214
MDS & MDT CODE(S):	MDS1005 & MDT 2004, 2011

PRODUCT SPECIFICATIONS

MATERIAL	Natural High Grade Rubber Latex
TYPE	Latex Surgical Gloves, Coated interior with Polymer
DESIGN & FEATURES	Hand specific, curved fingers, beaded cuff
COLOUR	White
STORAGE	The gloves shall maintain their properties when stored in a dry condition at temperature not higher than 30°C. Avoid direct sunlight
STERILIZATION	Yes (EO Sterilization)
SHELF LIFE	3 years from date of manufacturing
AQL (FACTORY STANDARD)	0.65
PACKING	<ul style="list-style-type: none">One pair in paper wallet and one wallet per pouch50 pouches in dispenser box8 boxes in one carton / As per customer requirement
APPLICATION	<ul style="list-style-type: none">General Surgery

KEY FUNCTIONAL BENEFITS

- Extra strength provides additional protection and strong tear resistance
- Curved fingers reduce hand fatigue while the beaded cuffs prevent roll-back and make gloves easy to put on
- Polymer coating interior enables incredible smooth donning
- Low levels of protein, offers excellent skin compatibility and eliminates risk of Latex Allergy
- Their anatomical shape, elasticity and softness provide superior fit, comfort, dexterity & flexibility

QUALITY STANDARDS

- ASTM:D-3577-2019 & EN455-1:2020, EN455-2:2015, EN 455-3:2015 & ISO10282:2014
- CE Mark – MDD Class IIa and IS 13422
- Manufactured under ISO 13485:2016 quality management system and ISO 9001:2015
- Using US FDA approved absorbable USP Grade corn starch
- Bioburden and sterility tested
- Manufactured in clean environment as per ISO14001:2015

QUALITY CONTROL

- 100% Inspected
- Factory Standard of 0.65 AQL for Pinholes

PHYSICAL DIMENSIONS

	Size	DR.GLOVE	ASTM D 3577 / IS 13422	EN 445
LENGTH (MM)		Min 280	Min 265	Min 280
WIDTH (MM)	5.5	70 ± 3	70 ± 6	72 ± 4
	6	76 ± 3	76 ± 6	77 ± 5
	6.5	83 ± 3	83 ± 6	83 ± 5
	7	89 ± 3	89 ± 6	89 ± 5
	7.5	95 ± 3	95 ± 6	95 ± 5
	8	102 ± 3	102 ± 6	102 ± 6
	8.5	108 ± 3	108 ± 6	108 ± 6
	9	114 ± 3	114 ± 6	114 ± 6
THICKNESS (MM)	Finger	0.16 Min	0.10 Min	0.10 Min
	Palm	0.16 Min	0.10 Min	0.10 Min
	Cuff	0.13 Min	0.10 Min	0.10 Min

PHYSICAL PROPERTIES

	DR.GLOVE	ASTM D 3577 / IS 13422	EN 445	
Tensile Strength (MPa)	BEFORE AGEING	Min 24	Min 24.0	9 N
Elongation At Break (%)		Min 800	Min 750	N/A
Tensile Strength (MPa)	AFTER AGEING	Min 20	Min 18.0	9 N
Elongation At Break (%)		Min 750	Min 560	N/A
Water Tightness Test G-1		AQL 0.65	AQL 0.65	AQL 0.65
Protein Content		< 200 µg/ dm ²	< 200 µg/ dm ²	< 200 µg/ dm ²

PRODUCT SAFETY CONFIRMANCE

BIOCOMPATIBILITY TESTED Test for Skin Irritation and Skin Sensitization	Pass
STERILITY TESTED To check the presence of bacterial in gloves after sterilization process	Pass
CHEMICAL RESIDUE TESTED To ensure the gloves are free from any chemical substances and safe to use	Pass
BIOBURDEN TESTED To test the numbers of bacterial living on the glove's surface that has not been sterilized	Pass
BACTERIAL ENDOTOXIN TESTED To check the presence of bacterial in gloves	Pass